



TEST REPORT TUCHENG

Date: Jun. 16, 2020 Date of Receipt: May. 28, 2020

Report No.: TFF9E595 Quantity: 1PC Page Order/Pages: (P1/6) Ref. No.: NIL


Report Title: EMS Components Assembly Inc(PH004) Item: Mask

Address: 117-A Technology Avenue, Laguna Technopark, Inc. SEPZ, Binan, Laguna, 4024 Philippines

Test Items		Test Results	Test Methods
Air Exchange Pressure (Pa/cm ²)	1	15.7	EN 14683:2019 Annex C
	2	15.7	
	3	16.7	
	4	16.7	
	5	18.6	

Note: Sample description is given by the client: Surgical Mask N88

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.



- Note:
- 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
 - 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
 - 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung Kao
Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839